

19L-SERVICE REGISTER

AUTO CLAVE

MODEL: Vertical "INLAB"
PURCHASE DATE: 09.07.97

SNO	Date of failure	Description of the failure	Date of the Complaint	Date of Service	RePair Details
1.	18.01.11	Not working	06.08.14	12.02.14. ✓	Heater element
2.	08.10.18	Not working	18.10.18	02.11.18	2000w Enamel heater - 1 no Pressure gauge meter - 1 no AC Power Porceline Plug - 1 no Rubber gasket - 1 no Vaccum releage brass valve - 1 pcs
3	08.10.19	Not working	21.10.19	01.11.19	Autoclave Heater Service

SUPPLIER NAME: Ponmani & Co

BILL DETAILS: 283/97-98


RATE: 19,178.00

NO: 1

AMOUNT charged	Service Eng sign	H.O.D sign	Remarks
₹ 1200/- B.M.No: 1176 12/02/14	<i>[Signature]</i> M. Sankar Kumar	<i>[Signature]</i> M. Sankar Kumar	# works well 29
Qty = 1 ₹ = 2600/- Bill no - 251/18-19	<i>[Signature]</i> (K. Loganathan)	<i>[Signature]</i> G. Ramesh	Autoclave works well.
Qty = 1 ₹ = 480.00	<i>[Signature]</i> (K. Loganathan)	<i>[Signature]</i> G. Ramesh	Working well.

Dr. G. MANIGANDAN, Ph.D., SET, NET
Head, Department of Biotechnology
J.J. College of Arts and Science (Autonomous)
Pudukkottai - 622 422

Dr. J. PARASURAMAN, M.A., M.B.A., M.C.A.,
M.PHIL., B.ED., P.E.D.
PRINCIPAL
J.J. College of Arts and Science
(Autonomous)
J.J. Nagar, Sivapuram Post,
PUDUKKOTTAI - 622 422.



① CUSTOMER LAMINAR AIR FLOW CHAMBER. (Horizontal)

MODEL :- CLEANAIR ATLANTIS

PURCHASE DATE: 16.7.97

S.NO	Date of failure	Description of the Complaint	Date of the Complaint	Date of Service	Repair details
1	10.02.17	Atlantis	16.02.17	31/03/17	general service HEPA clean
	10.02.17	cleanair	16.02.17	31/03/17	general service HEPA clean
2	05.10.18	NOT working	18.10.18	02.11.18	Acrylic door set replace -2set Manometer oil-2set

SUPPLIER NAME:- Ponmani & Co

BILL DETAILS:- 297/97-98


RATE : 52,900.00

NO: 21

Service details	Amount Charged	Service Engg sign	H.O.D sign	REMARKS
service & clean	₹. 600/- Bill No: 1516 Dt = 31/03/17	Septh N. Sathidkum Ennar Instent	G.P. Ramesh 31/03/17	HEPA filters are serviced
service & clean	₹. 600/- Bill No: 1516 Dt: 31/03/17	Septh N. Sathidkum Ennar Instent	G.P. Ramesh 31/03/17	HEPA filters are serviced
door set replace -2set Manometer oil-2set	Qty = 1 ₹ 6000/- Bill No: 251/ 18-19	Uthmanya (Ultra Scientific)	G.P. Ramesh 02/11/18	Condition is good.

Dr. G. MANIGANDAN, Ph.D., SET, NET
 Head, Department of Biotechnology
 J.J. College of Arts and Science (Autonomous)
 Pudukkottai - 622 422

Dr. J. PARASURAMAN, M.A., M.R., M.C.A.,
 M.PHIL, B.ED., P.T.D.
 PRINCIPAL
 J.J. College of Arts and Science
 (Autonomous)
 J.J. Nagar, Sivapuram Post,
 PUDUKKOTTAI - 622 422.



CYCLE MIXER

MODEL:- 'Remi' Make cat.No.-CM-101

PURCHASE DATE:- 29.11.99

S.NO	Date of Failure	Description of the Failure	Date of Complaint	Date of the Service
1	08.10.18	Not working	18.10.18	02.11.18
2	08.10.19	Not working	21.10.19	01.11.19

SUPPLIER NAME:- PONNANI & CO

105


BILL DETAILS:- 727/99-2000

RATE:- 1793/-

Repair Details	Amount charged	Service Eng H.O.D Sign	Remarks
Speed regulator board service	Qty = 2 = 2 x 600 = 1200/- Bill No:- 251/ 18-19	<i>[Signature]</i> G.P. [Signature] (Ultra Scientific)	Regulator is serviced
Vortex mixer Base Vacuum Push replace.	Qty = 4 = 4 x 140 = 560.00	<i>[Signature]</i> G.P. [Signature] (K. Jayanthi)	It was again working

[Signature]
Dr. G. MANIGANDAN, Ph.D. SET, NET
Head, Department of Biotechnology
J.J. College of Arts and Science (Autonomous)
Pudukkottai - 622 422

[Signature]
Dr. J. PARASURAMAN, M.A., M.B.A., M.PHIL., B. Ed.
PRINCIPAL
J.J. College of Arts and Science
(Autonomous)
J.J. Nagar, Sivapuram Post,
PUDUKKOTTAI - 622 422.



B-5

**Instrument Maintenance &
Service Register**

S. NO	Instruments	Page
1.	Rotary shaking machine	1
2.	Microscope	3
3.	Laminar flow chamber	5
4.	Cooling Centrifuge	7
5.	Micro Centrifuge	9
6.	Incubator Universal	11
7.	Colony Counter	13
8.	Cooling Centrifuge (M. Phil) lab	15
9.	Vertical Deep Freezer	17
10.	Water bath	18
11.	Photo electric Colorimeter	20
12.	clinical Centrifuge	22
13.	Auto clave	24
14.	Laminar flow chamber	26
15.	BOD Incubator	28
16.	High Speed Centrifuge RBC	30

Supplier details : Ponmani & Co

(6) 19

int	Date of Service	Repair Details	Amount charged	Servicer's Sig	HOD Sig
	18.08.14	Boothside 1/4 HP 1800 rpm motor replaced new one.	₹. 5500/- Bill no: 1175/18.08.14.	S. Senthil Kumar	P. Jeyaraj
	08.08.18	1. Switch - 3 nos 2. Manometer	₹. 1300/- Bill no: 190 18.19. Ultra Scientific	Whommy	P. Jeyaraj sb

**LB.4. INSTRUMENT-MAINTENANCE
AND SERVICE REGISTER**

PG and RESEARCH DEPARTMENT OF
BOTANY.

INSTRUMENT MAINTENANCE
AND SERVICE REGISTER.

S.No.	Name of the Instrument	Page.No.
1.	Digital Top pan Balance.	1 ✓
2.	Dissecting microscope.	3 ✓
3.	Microtome	5 ✓
4.	Laminar Air Flow ATH	7 ✓
5.	Laminar Air Flow ATV	9 ✓
6.	Spectrophotometer.	11 ✓
7.	Clinical Centrifuge	13 ✓
8.	Tissue Culture Rack	15 ✓
9.	AC UNIT I (Tissue Culture room)	17
10.	AC UNIT II (Inoculation room)	18
11.	Bacteriological Incubator	19
12.	Bacteriological oven	20
13.	Microscope	21
14.	Camera Lucida (Prism Type)	22
15.	Pressure Cooker	23
16.	Gas Stove	24
17.	pH Meter Pen Type.	25
18.	Haemocytometer.	26
19.	Spinix - vortex shaker	27
20.	Olympus Microscope	28
23.	Stereoscopic Microscope	29
24.	Refrigerator	30
25.	Bunsen Burner For Chamber.	31

MICROPHONE

S. No.	Date of failure	Nature of Failure	Date of Complaint	Date of Servicing	Service Engineer Signature	Remarks
1	21/11/16	Corrosion	21/11/16	21/11/16	<i>[Signature]</i> Senthil Kannan Coimbatore	n.ky
2	06/12/17	Corrosion	06/12/17	06/12/17	<i>[Signature]</i> C. Jayaraj (K. Loyd)	<i>[Signature]</i>
3	17/7/19	Knife	17/7/19	19/7/19	<i>[Signature]</i> C. Jayaraj (K. Loyd)	<i>[Signature]</i>

Laminar Air Flow ATH

S.No.	Date of Failure	Nature of Failures	Date of Complaint	Date of Servicing	Service Engineer Signature (K. J. A.)	Remarks
1.	20/8/19	Filter	20/8/19	23/8/19	C. J. A. (K. J. A.)	10/10
2.	04/02/20	Power switch	04/02/20	07/02/2020	C. J. A. (K. J. A.)	10/10

Laminar Air Flow ATV

S.No.	Date of Failure	Nature of Failure	Date of Complaint	Date of Servicing	Service Engineer Signature	Remarks
1	26/8/19	Filter	26/8/19	28/8/19	C. Ann (K. Jayshree)	
2	06/01/2020	UV Light	06/01/2020	06/01/2020	C. Ann (K. Jayshree)	
3	02/03/2020	Filter	02/03/2020	04/03/2020	C. Ann (K. Jayshree)	

Spectrophotometer

S.No.	Date of Failure	Nature of Failure	Date of Complaint	Services Date of Servicing	Service Engineer Signature	Remarks
1	27/11/11	Photo cell failure	2/12/11		C. Anur (K. Logesh)	
2	18/6/19	Digital Panel	19/6/19	21/6/19	C. Anur (K. Logesh)	
3	18/3/200	Lamp	19/3/200	20/3/2020	C. Anur (K. Logesh)	

CLINICAL CENTRIFUGE

S.No	Date & Fault Location	Nature & Failures	Date & Complant	Date & Servicing	Service Engineer Signature	Remarks
1	27/08/06	movement is Soly running Corrosion	31/08/06	21/11/06	Saf (Senthil Ponmani & Co Trichy) (Rs. 200)	found very H. K. J.
2	28/7/07	Centrifuge instrument is not working	10/9/07	Centrifuge Motor-ATUG 1390] is Sent through the Person Mr. Senthilkumar from Ponmani & Co, Trichy.	Saf (Senthilkumar) Ennarrinamb. Rs. 800/-	re-winding Dt: 13.3.08 H. K. J. 15/10/08
3	20/7/11	Centrifuge instrument is not working ATUG-1390.		motor taken for re-winding 27/02/12 motor returned 02/03/12 Bill no - 923 02/03/12 Bearing Set Rs. 800/-	Saf N. Senthilkumar Ennarrinamb. Saf 888 mark	
4	13/7/15	Rotor	17/7/15	17/7/15	C. Ann (K. Jayaram)	M. P. S.
5	28/2/17	Power cord	1/2/17	3/2/17	C. Ann (K. Jayaram)	M. P. S.
6	22/01/19	Centrifuge Service	22/01/19	22/01/19	C. Ann (K. Jayaram)	M. P. S.

TISSUE CULTURE RACK (with Automatic
Timer)

15

S. No	Date of Failure	Nature of Failure	Date of Complaint	Date of Servicing	Service Engineer Signature	Remarks
1	27/01/06	Timer is not working	27/01/06	18/02/06	[Signature]	Timer setting.
2	19/08/06	Timer failure	19/08/06	20/08/06	[Signature]	Timer fixing M. Jay
3	06/08/2018	Timer	06/08/2018	07/08/2018	C. Anurag (K. Jayaram)	[Signature]
4	27/01/2020	Lights	27/01/2020	27/01/2020	C. Anurag (K. Jayaram)	[Signature]

A/C UNIT I (TISSUE CULTURE ROOM) 17

S.No	Date of Jailer	Nature of Jailer	Date of Complaint	Date of Servicing	Service Engineer Signature	Remarks
1	28/07/06	AC - is Servicing	28/07/06	28/07/06	Bhasath J. M.	M. Jay
2	08/12/06	AC is service	08/12/06	08/12/06	Bhasath ENGINEER (Madurai)	G. VISWA
3	29/3/07	AC is Serviced	29/3/07	29/3/07	Bhasath ENGINEER (Madurai)	G. VISWA
4	29/11/07	AC is Serviced	29/11/07	29/11/07	P. ALY Bhasath ENGINEER MADURAI	
5	22/05/08	AC is Serviced	22/05/08	22/05/08	Dehesh	
6	19/11/08	"	19/11/08	19/11/08	D. Prabhu (Bharath Electricals)	
7	23.01.09	"	-	23.01.09	A. JERALD Bharath Engineers.	
8	25.07.09	"	-	25.01.09	Bharath Engineer MITHRANA.	
9	25-11-09	A/C SERVICE	25-11-09	25-11-09	R. Gobinath	
10	21/01/10	AC service		20/01/10	G. Nishant	
11	04/08/11	AC service		04/08/11	G. Nishant	
12	31/1/14	AC service		31/1/14	R. Ravi	
13	19/09/14	AC service		19/09/14	G. Nishant → Bharath	
14	13/07/15	AC service		13/07/15	G. Nishant → Bharath	
15	04/01/16	AC service		04/01/16	M. Thang → Bharath	
16	03/01/18	AC Service		03/01/18	G. Nishant → Bharath	
17	09/02/19	"		09/02/19	R. Ravi	
18	10/02/2020	"		10/02/2020	A. JERALD (Bharath Electronics)	



AC UNIT II (INOCULATION ROOM)

S.No	Date of failure	Nature of fault	Date of Complaint	Date of Servicing	Service Engineer signature	Remark
1	28/07/06	A/c is Servicing	28/07/06	28/07/06	Bharath	
2	08/12/06	A/c is serviced	08/12/06	08/12/06	G. V. S. S. G. Bharath engineer (Madurai)	
3	29/3/07	A/c is serviced	29/3/07	29/3/07	Bharath engineer (Madurai)	
4	29/11/07	"	29/11/07	29/11/07	Bharath engineer. Madurai	P.A.
5	22/05/08	"	22/05/08	22/05/08	D. Anand	
6	19/11/08	"	19/11/08	19/11/08	D. Prabhakaran (Bharath Electricals)	
7	23.01.09	"	--	23.01.09	A. Anand Bharath Engineers A-JEROLD Bharath	
8	25-11-09	"	"	"	R. Gobalakrishnan	
9	2/10/10	"	"	2/10/10	Bharath engineer G. V. S. S. G.	
10	04/08/11	"	"	04/08/11	Bharath engineer G. V. S. S. G.	
11	31/1/14	"	"	31/1/14	Bharath engineer R. Ravi	
12	19/09/14	A/c is Serviced.	19/09/14	19/09/14	Bharath engineer G. V. S. S. G.	
13	13/7/15	AC Service		13/7/15	Bharath engineer G. V. S. S. G.	
14	02/01/16	A/c Service		02/01/16	G. V. S. S. G.	
15	09.07.18	"			A. Anand	
16	09/02/19	"		09/02/19	R. Ravi	
17	10/02/2020	"		10/02/2020	A. Anand	

Microscope

S.No	Date of Failer	Nature of Failer	Date of Complaint	Date of Servicing	Service Engineer Signature	Remark
1	26/10/06	microscope set problem	26/10/06	6/10/06	S. Senthil (Senthil Ponmani & Co Trichy)	general M. L. V.
2	2/02/12	General Service		02/03/12	S. Senthil Senthil Kumar Emmarruthy exct 7.80x50 = 40.0%	88 man
					Bill no: 923 02/03/12.	
3	07/7/15	Service		07/7/15	S. Senthil (Senthil) Ponmani & Co	M. L. V.
4	18-2-16	general service		18-2-16	C. Suman (K. Suman)	M. L. V.
5	10/01/17	Service		10/01/17	C. Suman (K. Suman)	M. L. V.
6	27/06/18	Service		27/06/18	C. Suman (K. Suman)	M. L. V.
7	04/03/19	General Service		04/03/19	C. Suman (K. Suman)	M. L. V.
8	14/12/2020	Service		14/12/2020	C. Suman (K. Suman)	M. L. V.

Gas stove

S. No	Date of Failer	Nature of Failer	Date of Complant	Date of Servicing	Service Engineer Signature	Remarks
1.	27/12/07	Tube broken (2 nos)	27/12/07	27/12/07		New tube fixed Rs 20/- charged.
2.	03.12.18	Jet	03.12.18	03.12.18	Srimurugesan Gas service	
3.	10.02.20	Service	10.02.20	10.02.20	Srimurugesan Gas service	

Harencyrometer

Date S.No	Date of Failure	Nature of Failure	Date of Complaint	Date of Servicing	Service Engineer Signature	Remarks
--------------	--------------------	----------------------	----------------------	----------------------	----------------------------------	---------

1.	20.12.19	20.12.19 Service		20.12.19	C. Anura (K. Anura)	
----	----------	--------------------------------	--	----------	------------------------	--

Bussen Burner for Chamber

S.No	Date of failure	Nature of failure	Nature of Complaint	Date of Servicing	Service Engineer Signature	Remarks
1	10/9/07	Burner tube destroyed for long duration of usage	Informed to Mr. Sander PDKT	12/9/07 (New tube fixed)	S. Sundarajan	
2	18/07/2016	Burner full service	18/07/2016	20/07/2016	Srinivasan Gas Service Centre	

Dr. M. Arumugam, M.Sc., M.Phil., Ph.D.
 Assistant Professor & Head
 PG and Research Dept. of Botany
 J.J. College of Arts and Science
 Sivapuram, Pudukkottai - 622 422
 Tamil Nadu, India.



Dr. J. PARASURAMAN, M.A., M.B.A., M.C.A.,
 M.Phil., B.Ed., Ph.D.
 PRINCIPAL
 J.J. College of Arts and Science
 (Autonomous)
 J.J. Nagar, Sivapuram Post,
 PUDUKKOTTAI-622 422

38

**INSTRUMENTS MAINTENANCE
AND SERVICE REGISTER**

Sl. No.	Name of the Equipments	Qty	Whom sent to the party	Return date	Remarks
21/8	<u>IE Lab.</u>	2 ✓			
1.	5V DC Regulated Power supply (micronet 5512 make)	1 ✓	Kumaras	One week	
2	Dual Power Supply 2A (kss make)	1 ✓	Scientific Systems		
3	Microprocessor 8085 kit (model: Micro 85L)	1 ✓	Tricky		
4	Microprocessor kit 8085 model: Micro 85F. Slms. 917, 918	2 ✓	A. S. Sathyanarayanan		
5	Microprocessors kit 8086 Model: Micro-85FD Slms. 2650, 2689, 2687, 2685	4 ✓	None	2/4/18	Microprocessor 8085 slms 2650 without Servicing
6	Afo (Testronix mds) Slms. 950806	9 ✓			RS. 7897/- dt 24/5/96
	021014 950818				Afo slms.
	950807 021028				961330 without Servicing
	021023 961233				RS. 21850
	021018 961235				dt 31/7/2002
7	DPT switch	2 ✓			
		21	sk		
	<u>Phy Lab.</u>				
1.	Battery Eliminator (Pico make) (0-12v)	3 ✓			
2	Spectrometer (Weswox Optik make)	2 ✓		2/4/18	
3	Spectrometer transformer	2 ✓			
4	Travelling Microscope (model weswox)	3 ✓			1 no without Servicing
5	Telescope (model weswox)	2 ✓			
		17	✓ ✓ ✓		

Date	Name of the Equipment	Qty	whom sent to the party	Return date	Remarks
6	Stop clock, CFSail meter	3	Kumarav Scientific		
7	four probe Power supply	2	System Tridhey		
8	four probe oven	3	M. Sridhar Kumar		Phy Ino. no 540/2017-18 dt 28/3/18 and
		20			
	Multimeter (Master W MA2A)	1		2/4/18 Pd	Phy: Ino. no 539/201718 dt 28/3/2018
	Plug Computer	2		RS. 16,900/2	
		23			

20. Physics Dept:- 20/2/19

- 1 Spectrometer - Weswox 5 ✓ Ultra Scientific
- 2 Travelling Microscope - 3 ✓ Supplies
- 3 Gauss Meter Digital - Pico 2 ✓ Pudukkottai
- 4 Analog desk top meter - Osan 20 ✓
- 5 Resistance Box Adial Pico 06 ✓
- 6 four Probe heater Assembly - Besto 01 ✓
- 7 Reading telescope - Weswox 4 ✓
- 8 Rheostat - Besto 2 ✓
- 9 Standard Resistance Round type - Besto 2 ✓
- 10 Dimmer Stat - Besto 01 ✓
- 11 Cross wire eye piece 10 x Ramsden type - Weswox 08 ✓

Amount
Rs. 17300/-

Received
the items
Skt
27/4/19

Dr. M.K MURALI, M.Sc., M.Phil., Ph.D.,
HEAD AND ASSISTANT PROFESSOR,
P.G. AND RESEARCH DEPARTMENT OF PHYSICS
J.J. COLLEGE OF ARTS AND SCIENCE (AU),
J.J. NAGAR, SIVAPURAM,
PUDUKKOTTAI - 622 422.




Dr. J. PARASURAMAN, M.A., M.Ed., M.C.A.,
M.Phil., B.Ed., Ph.D.,
PRINCIPAL
J.J. COLLEGE OF ARTS AND SCIENCE
(Autonomous)
J.J. Nagar, Sivapuram Post,
PUDUKKOTTAI - 622 422

PGT-LAB

STOCK

REGISTER






J.J. COLLEGE OF ARTS AND SCIENCE
 J.J. NAGAR, SIVAPURAM, PUDUKKOTTAI - 622 422.

STOCK REGISTER

Article : *Work station - 1*

S. No.	Date	Description	Supplier Name and Bill Details	Rate	Quantity Received	Issue	Stock in Hand	Remarks
1.	12/5/14	13 - Intel	CHALLENGER					
		Asrock Mother Board	COMPUTER					
		2 GIB DDR3 RAM	CHENNAI					
		80 GIB HDD		17,500	1		1	
		Acer 15" Monitor	Bill No: 33247					
		Zebtronc Keyboard	Bill Date					
		Zebtronc Mouse	7/5/14					
		Zebtronc ATX Cabinet	Bill Amount					
			10,50,000					
			verified by					
								M. Vanitha 11/4/15 (Dr. M. Vanitha)
		13/11/15 New Logitech Key Board	Installed.					
			<i>D. Suresh</i>					
			Verified by					A. Suresh 07/04/16 CA Mania Stella
			Transferred to BHM Net Lab					M. Suresh 12/4/17

Article : System - 2

S. No.	Date	Description	Supplier Name and Bill Details	Rate	Quantity Received	Issue	Stock in Hand	Remarks
		Intel i3	CHALLENGER					
		Asrock Mother Board	COMPUTER					
		2 GB DDR3 RAM	CHENNAI					
		80 GB sata HDD		17,500	1		1	
		Acer 15" LCD monitor	Bill NO - 33247					
		Zebtronc key Board	Bill Date					
		Zebtronc Mouse	7/5/2014					
		Zebtronc ATX Cabinet	Bill Amount					
			10,50,000					
		Verified by	M. 					
			(Cdr. M. Varnitha)					
		Verified by	A.  07/04/16					
			(A. Maria Stella)					
		Verified by	M.  12/4/17					



J.J. COLLEGE OF ARTS AND SCIENCE

J.J. NAGAR, SIVAPURAM, PUDUKKOTTAI - 622 422.




STOCK REGISTER

84



Article : Workstation - 81

Article

S. No.	Date	Description	Supplier Name and Bill Details	Rate	Quantity Received	Issue	Stock in Hand	Remarks	S. No.	
1.		Pentium - 4 Processor			1		1			
		Samsung CRT monitor								
		512 MB DDR1 RAM								
		80 GB IDE H DD								
		Logitech Key Board								
		Logitech Mouse								
		Zebra Cabinet								
		Verified by A. Jey (A. Maria Stella) on 10/04/16								
		Verified by H. Jey 12/6/17								
		 								
		Dr. S. SUDHA M.Sc., M.Phil., Ph.D., SET Vice Principal and Head, Department of Computer Science J.J. College of Arts and Science Autonomous Pudukkottai - 622 422								
						Dr. J. PARASURAMAN , M.A., B.A., B.L.S., B.Ed., M.Ed., Ph.D., SET PRINCIPAL J.J. College of Arts and Science (Autonomous) J.J. Nagar, Sivapuram Post, PUDUKKOTTAI-622 422				

PG Lab

STOCK

REGISTER



164



J.J. COLLEGE OF ARTS AND SCIENCE

(AUTONOMOUS)

J.J. NAGAR, SIVAPURAM, PUDUKKOTTAI - 622 422.

Article : *System 100*

STOCK REGISTER

S. No.	Date	Description	Supplier Name and Bill Details	Rate	Quantity Received	Issue	Stock in Hand	Remarks
	Jun 2018	Intel Core i3-3220, 3.20 GHz Processor, ASUS H81 Mother board, DDR3 2GB RAM, 5TB 6TB Harddisk, Zebronics mouse, Zebronics keyboard, Acer 15.6 monitor, Zebronics Cabinet.			1			
	27/4/19	Verified by M. Nithya Kalyani Meledu						27/4/19
	 Dr. S. SUDHA M.Sc., M.Phil., Ph.D., B.Ed. Vice Principal and Head, Department of Computer Science J.J. College of Arts and Science (Autonomous) Pudukkottai - 622 422		 Dr. J. PARASURAMAN, M.A., M.B.A., M.C.A., M.Phil., B.Ed., Ph.D. PRINCIPAL J.J. College of Arts and Science (Autonomous) J.J. Nagar, Sivapuram Post, PUDUKKOTTAI - 622 422					
		